NMS Labs
200 Welsh Road, Horsham, PA 19044-2208
Phone: (215) 657-4900 Fax: (215) 657-2972
e-mail: nms@nmslabs.com
Robert A. Middleberg, PhD, F-ABFT, DABCC-TC, Laboratory

CONFIDENTIAL

Demo Report

Report Issued 04/03/2020 11:22

To: **88888**
Forensic Example Report
Attn: Example Reports
200 Welsh Road
Horsham, PA 19044

Patient Name 0939B
Patient ID 0939B
Chain 20000748
Age Not Given **DOB** Not Given
Gender Not Given
Workorder 20000748

Positive Findings:

None Detected

See Detailed Findings section for additional information

Testing Requested:

Analysis Code	Description
0939B	Calcium - Total, Postmortem, Blood (Forensic)

Specimens Received:

ID	Tube/Container	Volume/ Mass	Collection Date/Time	Matrix Source	Miscellaneous Information
001	Clear vial	Not Given	Not Given	Blood	

All sample volumes/weights are approximations.

Specimens received on 03/31/2020.

Detailed Findings:

Examination of the specimen(s) submitted did not reveal any positive findings of toxicological significance by procedures outlined in the accompanying Analysis Summary.

Analysis Summary and Reporting Limits:

All of the following tests were performed for this case. For each test, the compounds listed were included in the scope. The Reporting Limit listed for each compound represents the lowest concentration of the compound that will be reported as being positive. If the compound is listed as None Detected, it is not present above the Reporting Limit. Please refer to the Positive Findings section of the report for those compounds that were identified as being present.

Acode 0939B - Calcium - Total, Postmortem, Blood (Forensic)

-Analysis by Inductively Coupled Plasma/Optical Emission Spectrometry (ICP/OES) for:

Compound	Rpt. Limit	Compound	Rpt. Limit
Calcium	0.50 mg/dL		

Specimens for elemental testing should be collected in certified metal-free containers. Elevated results for elemental testing may be caused by environmental contamination at the time of specimen collection and should be interpreted accordingly. It is recommended that unexpected elevated results be verified by testing another specimen.