



NMS Labs

CONFIDENTIAL

200 Welsh Road, Horsham, PA 19044-2208
Phone: (215) 657-4900 Fax: (215) 657-2972
e-mail: nms@nmslabs.com

Robert A. Middleberg, PhD, F-ABFT, DABCC-TC, Laboratory Director

Demo Report

Report Issued 03/30/2020 12:39
Last Report Issued 06/09/2014 07:41

88888
Clinical Example Report
Attn: Example Reports
200 Welsh Road
Horsham, PA 19044

Patient Name 2570R
Patient ID 2570R
Chain 14001435
Age Not Given **DOB** Not Given
Gender Not Given
Workorder 14001435
Received 06/05/2014 06:57

Sample ID 14001435-001
Matrix RBCs
Patient Name 2570R
Patient ID 2570R
Container Type Clear vial

Collect Dt/Tm Not Given
Source Not Given

Approx Vol/Weight Not Given

Receipt Notes None Entered

Analysis and Comments	Result	Units	Reporting Limit	Notes
-----------------------	--------	-------	-----------------	-------

2570R Manganese, RBCs

Analysis by Inductively Coupled Plasma/Mass Spectrometry (ICP/MS)

Manganese	None Detected	mcg/L	11	
-----------	---------------	-------	----	--

NMS Labs derived data:
2.5th - 97.5th percentile range is
4.7 to 20 mcg/L (n=2022).

The RBC sample used for analysis was measured by weight and multiplied by the density of human RBC (1.10 g/mL) to obtain mcg/L units.

Specimens for elemental testing should be collected in certified metal-free containers. Elevated results for elemental testing may be caused by environmental contamination at the time of specimen collection and should be interpreted accordingly. It is recommended that unexpected elevated results be verified by testing another specimen.

This test was developed and its performance characteristics determined by NMS Labs. It has not been cleared or approved by the US Food and Drug Administration.