



NMS Labs

CONFIDENTIAL

200 Welsh Road, Horsham, PA 19044-2208
Phone: (215) 657-4900 Fax: (215) 657-2972
e-mail: nms@nmslabs.com

Robert A. Middleberg, PhD, F-ABFT, DABCC-TC, Laboratory Director

Demo Report

Report Issued 03/30/2020 13:37
Last Report Issued 01/16/2013 08:53

88888
Clinical Example Report
Attn: Example Reports
200 Welsh Road
Horsham, PA 19044

Patient Name 4618B-POS
Patient ID 4618B-POS
Chain 13000178
Age Not Given **DOB** Not Given
Gender Not Given
Workorder 13000178
Received 01/15/2013 09:09

Sample ID 13000178-001
Matrix Blood
Patient Name 4618B-POS
Patient ID 4618B-POS
Container Type Clear vial

Collect Dt/Tm Not Given
Source Not Given

Approx Vol/Weight Not Given

Receipt Notes None Entered

Analysis and Comments	Result	Units	Reporting Limit	Notes
-----------------------	--------	-------	-----------------	-------

4618B Trichlorobenzenes, Blood

Analysis by Gas Chromatography (GC)

1,2,4-Trichlorobenzene	50	ng/mL	5.0	
Hexachlorobutadiene interferes with 1,2,4 - trichlorobenzene in this analysis. The presence of hexachlorobutadiene will adversely affect the quantitation of 1,2,4 - trichlorobenzene. If hexachlorobutadiene is a potential interferent in this case, call the laboratory for alternate quantitative procedures.				
1,2,3-Trichlorobenzene	50	ng/mL	5.0	
1,3,5-Trichlorobenzene	50	ng/mL	5.0	
No reference data available.				

This test was developed and its performance characteristics determined by NMS Labs. It has not been cleared or approved by the US Food and Drug Administration.