



NMS Labs

CONFIDENTIAL

200 Welsh Road, Horsham, PA 19044-2208
Phone: (215) 657-4900 Fax: (215) 657-2972
e-mail: nms@nmslabs.com

Robert A. Middleberg, PhD, F-ABFT, DABCC-TC, Laboratory Director

Demo Report

Report Issued 03/30/2020 13:40
Last Report Issued 01/28/2013 09:45

88888
Clinical Example Report
Attn: Example Reports
200 Welsh Road
Horsham, PA 19044

Patient Name 4730U
Patient ID 4730U
Chain 13000464
Age Not Given **DOB** Not Given
Gender Not Given
Workorder 13000464
Received 01/28/2013 08:00

Sample ID 13000464-001
Matrix Urine
Patient Name 4730U
Patient ID 4730U
Container Type Clear vial

Collect Dt/Tm Not Given
Source Not Given

Approx Vol/Weight Not Given

Receipt Notes None Entered

Analysis and Comments	Result	Units	Reporting Limit	Notes
-----------------------	--------	-------	-----------------	-------

4730U Tungsten, Urine

Analysis by Inductively Coupled Plasma/Mass Spectrometry (ICP/MS)

Tungsten Normally less than 1 mcg/L	None Detected	mcg/L	1.0	
--	---------------	-------	-----	--

Specimens for elemental testing should be collected in certified metal-free containers. Elevated results for elemental testing may be caused by environmental contamination at the time of specimen collection and should be interpreted accordingly. It is recommended that unexpected elevated results be verified by testing another specimen.

This test was developed and its performance characteristics determined by NMS Labs. It has not been cleared or approved by the US Food and Drug Administration.